



CALIBRATION RESULTS

MODEL	SB-4500	DATE	08/11/2025
SERIAL NUMBER	472915000012	QUANTITY	1pc
APPEARANCE	Clean, no corrosion, no burr and other mechanical damage		
MAJOR FUNCTION	Testing, calibration, setting functions are working properly.		
DETECTING GAS	Detecting range	Indication error	
C ₂ H ₅ OH	0~4.000 g/L	±0,048 g/L	
DETECTION POINT	0,488 g/L		
TEST CONDITION	Temperature: 23°C ± 2°C		
TEST CONDITION	Relative humidity: 20%RH~80%RH		
RESPONSE TIME	Qualified		
DISPLAY	Qualified		
TEST RESULT	Qualification		

www.safebreathalyzer.com

